

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/01/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:48
INTERNAL STANDARD	VERIFIED	18:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:51
BLANK TEST	0.00	18:52

