

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 02/16/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:54
INTERNAL STANDARD	VERIFIED	19:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:57
BLANK TEST	0.00	19:58

