

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 04/13/2022

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:17
INTERNAL STANDARD	VERIFIED	19:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:18
BLANK TEST	0.00	19:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:19
BLANK TEST	0.00	19:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:20
BLANK TEST	0.00	19:21

