

SUPERVISOR TEST

LEB

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 04/20/2022

OPERATOR NAME:
 ROBERT S CORNISH
 Permit #: 48068
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:33
INTERNAL STANDARD	VERIFIED	18:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:34
BLANK TEST	0.00	18:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:35
BLANK TEST	0.00	18:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:36
BLANK TEST	0.00	18:37

