

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LBB

Date: 04/27/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:09
INTERNAL STANDARD	VERIFIED	19:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:10
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:12
BLANK TEST	0.00	19:13

