

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 06/29/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20

