

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/01/2022

LEB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:11
INTERNAL STANDARD	VERIFIED	22:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:11
BLANK TEST	0.00	22:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:14
BLANK TEST	0.00	22:15

