

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/06/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:15
INTERNAL STANDARD	VERIFIED	06:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:15
BLANK TEST	0.00	06:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:16
BLANK TEST	0.00	06:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:17
BLANK TEST	0.00	06:18

