

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/13/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:48
INTERNAL STANDARD	VERIFIED	19:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:49
BLANK TEST	0.00	19:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:50
BLANK TEST	0.00	19:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:51
BLANK TEST	0.00	19:52

