

TECHNICIAN TEST

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 09/05/2023

Nominal = 0.100 g/210L

BLANK TEST	0.000	11:10
INTERNAL STANDARD	VERIFIED	11:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.099	11:10
BLANK TEST	0.000	11:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.098	11:11
BLANK TEST	0.000	11:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.099	11:13
BLANK TEST	0.000	11:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.099	11:14
BLANK TEST	0.000	11:15

Average = 0.0988
 Std Dev = 0.0005

