

SOLUTION CHANGE

LEB

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606

Date: 08/01/2023

OPERATOR NAME:  
 JASON L OSIKA  
 Permit #: 35567  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:27
INTERNAL STANDARD	VERIFIED	19:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:27
BLANK TEST	0.00	19:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:30
BLANK TEST	0.00	19:31

