

SOLUTION CHANGE

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606

LEB

Date: 11/01/2023

OPERATOR NAME:  
 JOHN P CUSACK  
 Permit #: 35563  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:26
INTERNAL STANDARD	VERIFIED	18:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:26
BLANK TEST	0.00	18:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:29
BLANK TEST	0.00	18:30

