

SOLUTION CHANGE

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

LEB

Date: 02/01/2023

OPERATOR NAME:
 JOHN P CUSACK
 Permit #: 35563
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:00
INTERNAL STANDARD	VERIFIED	02:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03

