

SOLUTION CHANGE

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/01/2023

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:00
INTERNAL STANDARD	VERIFIED	06:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:00
BLANK TEST	0.00	06:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:01
BLANK TEST	0.00	06:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:03
BLANK TEST	0.00	06:03

