

SOLUTION CHANGE

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 03/28/2023

OPERATOR NAME:
 JAY V SORRENTINO
 Permit #: 38962
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:02
INTERNAL STANDARD	VERIFIED	16:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:03
BLANK TEST	0.00	16:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:04
BLANK TEST	0.00	16:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:05
BLANK TEST	0.00	16:06

