

SOLUTION CHANGE

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 04/01/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:22
INTERNAL STANDARD	VERIFIED	17:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:23
BLANK TEST	0.00	17:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:24
BLANK TEST	0.00	17:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:25
BLANK TEST	0.00	17:26

