

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/27/2023

LEB

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:04
INTERNAL STANDARD	VERIFIED	18:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:05
BLANK TEST	0.00	18:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:06
BLANK TEST	0.00	18:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:07
BLANK TEST	0.00	18:08

