

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/12/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:57

