

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 05/24/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:21
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:25

