

SUPERVISOR TEST

ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606

LEB

Date: 06/08/2023

OPERATOR NAME:  
 STEVEN P EILERS  
 Permit #: 49321  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:51
INTERNAL STANDARD	VERIFIED	00:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:54
BLANK TEST	0.00	00:55

