

SUPERVISOR TEST

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

LEB

Date: 06/14/2023

OPERATOR NAME:
 JOHN P CUSACK
 Permit #: 35563
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:22
INTERNAL STANDARD	VERIFIED	19:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:25
BLANK TEST	0.00	19:26

