

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/05/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:36
INTERNAL STANDARD	VERIFIED	01:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40

