

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LER

Date: 07/19/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:49
INTERNAL STANDARD	VERIFIED	02:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:50
BLANK TEST	0.00	02:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:51
BLANK TEST	0.00	02:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:52
BLANK TEST	0.00	02:53

