

SUPERVISOR TEST

LE13

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/26/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:42
INTERNAL STANDARD	VERIFIED	21:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:43
BLANK TEST	0.00	21:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:44
BLANK TEST	0.00	21:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:45
BLANK TEST	0.00	21:46

