

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/06/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:57
INTERNAL STANDARD	VERIFIED	16:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:57
BLANK TEST	0.00	16:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:58
BLANK TEST	0.00	16:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:00
BLANK TEST	0.00	17:01

