

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/27/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:38
INTERNAL STANDARD	VERIFIED	04:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:38
BLANK TEST	0.00	04:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:40
BLANK TEST	0.00	04:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:41
BLANK TEST	0.00	04:42

