

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LETS

Date: 10/18/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:53
INTERNAL STANDARD	VERIFIED	22:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:53
BLANK TEST	0.00	22:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:54
BLANK TEST	0.00	22:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:56
BLANK TEST	0.00	22:56

