

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 01/05/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 22390
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14

