

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/11/2023

LEB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:22
INTERNAL STANDARD	VERIFIED	04:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:22
BLANK TEST	0.00	04:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:23
BLANK TEST	0.00	04:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:24
BLANK TEST	0.00	04:25

