

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/22/2023

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:21
INTERNAL STANDARD	VERIFIED	19:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25

