

SUPERVISOR TEST

LEB

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 03/22/2023

OPERATOR NAME:
 JOHN P CUSACK
 Permit #: 35563
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:07
INTERNAL STANDARD	VERIFIED	23:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:09
BLANK TEST	0.00	23:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:10
BLANK TEST	0.00	23:11

