

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 11/15/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:08
INTERNAL STANDARD	VERIFIED	19:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:09
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:10
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12

