

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/29/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:22
INTERNAL STANDARD	VERIFIED	00:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:24
BLANK TEST	0.00	00:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:25
BLANK TEST	0.00	00:26

