

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/06/2023

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:29
INTERNAL STANDARD	VERIFIED	04:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:29
BLANK TEST	0.00	04:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:32
BLANK TEST	0.00	04:33

