

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/20/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:34
INTERNAL STANDARD	VERIFIED	04:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:34
BLANK TEST	0.00	04:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:36
BLANK TEST	0.00	04:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:37
BLANK TEST	0.00	04:38

