

SOLUTION CHANGE

LETB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 01/31/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:25
INTERNAL STANDARD	VERIFIED	05:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:26
BLANK TEST	0.00	05:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:27
BLANK TEST	0.00	05:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:28
BLANK TEST	0.00	05:29

