

SOLUTION CHANGE

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/21/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:14
INTERNAL STANDARD	VERIFIED	23:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18

