

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LIB

Date: 05/01/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:29
INTERNAL STANDARD	VERIFIED	18:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:30
BLANK TEST	0.00	18:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:31
BLANK TEST	0.00	18:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:32
BLANK TEST	0.00	18:33

