

SOLUTION CHANGE

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 09/01/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:57
INTERNAL STANDARD	VERIFIED	17:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:59
BLANK TEST	0.00	18:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:00
BLANK TEST	0.00	18:01

