

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/01/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:22
INTERNAL STANDARD	VERIFIED	15:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:23
BLANK TEST	0.00	15:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:25
BLANK TEST	0.00	15:26

