

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 12/02/2024

OPERATOR NAME:  
BRADY A WOOD  
Permit #: 52004  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:22
INTERNAL STANDARD	VERIFIED	02:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25

LEB

