

SOLUTION CHANGE

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/01/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:07
INTERNAL STANDARD	VERIFIED	03:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:08
BLANK TEST	0.00	03:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:09
BLANK TEST	0.00	03:10

