

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/02/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:40
INTERNAL STANDARD	VERIFIED	20:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:41
BLANK TEST	0.00	20:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:42
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:43
BLANK TEST	0.00	20:44

