

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/03/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:39
INTERNAL STANDARD	VERIFIED	19:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:39
BLANK TEST	0.00	19:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:41
BLANK TEST	0.00	19:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:42
BLANK TEST	0.00	19:43

