

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 02/08/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:32
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36

