

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/13/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:24
INTERNAL STANDARD	VERIFIED	04:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:24
BLANK TEST	0.00	04:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:26
BLANK TEST	0.00	04:27

