

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/10/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:48
INTERNAL STANDARD	VERIFIED	04:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:48
BLANK TEST	0.00	04:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:49
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:50
BLANK TEST	0.00	04:51

