

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 04/24/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:24
INTERNAL STANDARD	VERIFIED	04:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:25
BLANK TEST	0.00	04:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:26
BLANK TEST	0.00	04:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:27
BLANK TEST	0.00	04:28

