

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 07/03/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:21
INTERNAL STANDARD	VERIFIED	05:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:22
BLANK TEST	0.00	05:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:23
BLANK TEST	0.00	05:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:24
BLANK TEST	0.00	05:25

