

SUPERVISOR TEST

LEB

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 07/17/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:01
INTERNAL STANDARD	VERIFIED	04:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04

